

MEASUREMENT SYSTEM ANALYSIS

Gage R&R Study (ANOVA Method) · AIAG MSA 4th Edition

My Metrology Laboratory

Study No:

GRR-2026-0001

Date: 2026-05-22

INSTRUMENT MIC-001	CHARACTERISTIC Length	PART Gauge Block 25mm
PARTS 5	OPERATORS 3 (Operator A, Operator B, Operator C)	TRIALS 2

ANOVA TABLE

Source	SS	df	MS
Parts	0.0032	4	0.0008
Operators	0.0000	2	0.0000
Part × Operator	0.0000	8	0.0000
Equipment (Repeat.)	0.0000	15	0.0000
Total	0.0032	29	

VARIANCE COMPONENTS & GAGE R&R

Source	Variance	Std Dev	% Contribution	% Study Var
Repeatability (EV)	8.333e-7	9.129e-4	0.62%	7.86%
Reproducibility (AV)	3.250e-7	5.701e-4	0.24%	4.91%
Total Gage R&R	1.158e-6	0.001076	0.86%	9.27%
Part-to-Part (PV)	1.337e-4	0.011563	99.14%	99.57%
Total Variation	1.349e-4	0.011613	100%	100%

% GAGE R&R (Study Variation):

9.27%

NDC:

15

% Tolerance:

12.92%

ACCEPTABLE

Measurement system is acceptable.

Acceptance criteria (AIAG MSA): %GRR < 10% acceptable; 10–30% conditionally acceptable; > 30% unacceptable. NDC ≥ 5 indicates the system can adequately distinguish parts. % Study Variation uses 6σ